

**Annex K6**  
**Phase IV Experiments**

This page intentionally left blank

**Table 1 Phase IV Agonist Range Finder Plates Tested at XDS**

Experiment I.D.	Date	Induction <sup>1</sup>	DMSO	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used
XPIIX0036X0033X0027X0025X0122X0123AGRF081309	13-Aug-09	2.56	1240	No	Failed Induction
XPIIX0036X0033X0027X0025X0122X0123AGRF081509	15-Aug-09	5.47	937	Yes	
XPIIX0065X0120X0119AGRF082109	21-Aug-09	3.52	1311	Yes	
XPIVX0108X0112X0128X0115X0109X0125AGRF082509	25-Aug-09	3.75	3021	Yes	
XPIVX0117X0130X0121X0120X0119X0114AGRF082509	25-Aug-09	6.78	1514	Yes	
XPIVX0118X0111X0107X0116X0127X0126AGRF082509	25-Aug-09	5.99	2172	Yes	
2009-11-08 XPIVAgRFX0131X0124X0113X0129X0110	18-Nov-09	3.45	4799	Yes	

<sup>1</sup> Induction for range finder plates is measured by dividing the averaged highest E2 reference standard RLU value by the averaged DMSO control RLU value.

**Table 2 Phase IV Antagonist Range Finder Plates Tested at XDS**

Experiment I.D.	Date	Reduction <sup>1</sup>	DMSO	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used
XPIIX0067X0147X0148X0144X0141X0138ANTRF081309	13-Aug-09	2.49	1746	No	Failed Reduction
XPIIX0067X0147X0148ANTRF081509	15-Aug-09	10.67	797	Yes	
2009-18-11-09XPIVANTRFX0140	18-Nov-09	4.53	3433	Yes	
2009-18-11XPIVANTRFX0155X0137X0149X0151X0152X0153	18-Nov-09	3.73	4007	Yes	
2009-11-18XPIVRFX0142X0143X0146X0156X0132X0136	18-Nov-09	4.00	2818	Yes	
2009-11-20XPIVANTRFX0139X0150X0135X0134X0133	20-Nov-09	6.19	2734	Yes	

<sup>1</sup> Reduction for range finder plates is measured by dividing the averaged highest Ral/E2 reference standard RLU value by the averaged DMSO control RLU value.

This page intentionally left blank

**Table 3 Phase IV Agonist Comprehensive Test Plates Tested at XDS**

Test Substance	Date	DMSO	Induction <sup>1</sup>	E2 Reference Standard EC <sub>50</sub> (µg/mL)	MET	Test Substance Dilution	Test Substance EC <sub>50</sub> (µg/mL)	Positive Using Threshold Method	Positive Using Revised Criteria	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used	Experiment I.D.
AMP	13-Nov-09	7138	4.61	2.45 x 10 <sup>-6</sup>	532	1:2	-	Negative	Negative	No	Failed MET	2009-11-13XPIVAgCTT X0114X0128
AMP	16-Nov-09	4899	3.69	2.48 x 10 <sup>-6</sup>	6663	1:2	-	Positive	Negative	Yes	Experimenter error	2009-11-16XPIVAgCTT X0065X0128
AMP	07-Dec-09	2386	3.92	2.68 x 10 <sup>-6</sup>	5849	1:2	-	Positive	Negative	Yes		2009-12-07XPIVAgCTX0 111X0128
APO	28-Dec-09	3675	3.10	3.55 x 10 <sup>-6</sup>	5756	1:5	-	Negative	Negative	Yes		2009-12-28XPIVAgCTT X0117X0127
APO	16-Nov-09	5277	3.43	1.24 x 10 <sup>-6</sup>	6378	1:5	-	Negative	Negative	Yes		2009-11-16XPIVAgCTT X0127X0130
APO	07-Dec-09	3541	2.91	1.55 x 10 <sup>-6</sup>	7065	1:5	-	Positive	Negative	No	Failed Induction	2009-12-07XPIVAgCTX0 117X0127
BENZ*	13-Dec-09	1810	3.52	2.13 x 10 <sup>-6</sup>	5202	1:5	1.16 x 10 <sup>-1</sup>	Positive	Inconclusive	Yes		2009-12-13XPIVAgCTT X0160X0116
BENZ	21-Jun-10	915	3.43	1.59 x 10 <sup>-5</sup>	5108	1:5	1.16 x 10 <sup>-1</sup>	Positive	Positive	Yes		2010-06-21X0129X0160 XPIVAgCTRe-Run
BICAL	18-Nov-09	2482	3.57	3.96 x 10 <sup>-6</sup>	5932	1:5	-	Negative	Negative	Yes	Experimenter error	2009-11-18 XPIVAgCTTX0 118X0126
BICAL	07-Dec-09	2482	4.33	2.80 x 10 <sup>-6</sup>	5868	1:5	-	Positive	Negative	Yes		2009-12-07XPIVAgCTX0 119X0126
CHX	13-Dec-09	1825	3.13	2.74 x 10 <sup>-6</sup>	6247	1:2	-	Positive	Negative	Yes		2009-12-13XPIVAgCTT X0113X0124

ICCVAM BG1Luc ER TA Evaluation Report

Test Substance	Date	DMSO	Induction <sup>1</sup>	E2 Reference Standard EC <sub>50</sub> (µg/mL)	MET	Test Substance Dilution	Test Substance EC <sub>50</sub> (µg/mL)	Positive Using Threshold Method	Positive Using Revised Criteria	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used	Experiment I.D.
CHY	13-Nov-09	4753	6.47	2.93 x 10 <sup>-6</sup>	2847	1:2	-	Positive	Inconclusive	Yes	Experimenter error	2009-11-13XPIVAgCTT X0125X0117
CHY	04-Dec-09	2187	3.93	3.87 x 10 <sup>-6</sup>	3761	1:2	-	Positive	Inconclusive	Yes		2009-12-04XPIVAgCTX0 112X0125
CHY	13-Dec-09	1488	4.02	1.95 x 10 <sup>-6</sup>	6076	1:2	8.14 x 10 <sup>-1</sup>	Positive	Positive	Yes		2009-12-13XPIVAgCTT X0131X0150
CYP	12-Nov-09	3889	9.35	4.00 x 10 <sup>-5</sup>	3191	1:5	-	Positive	Negative	Yes	Experimenter error	2009-11-12XPIVAgCTT X0123X0122
CYP	07-Dec-09	2723	4.00	2.00 x 10 <sup>-6</sup>	7034	1:5	-	Positive	Inconclusive	Yes		2009-12-07XPIVAgCTX0 123X0122
CYP	24-Aug-09	2135	3.15	3.26 x 10 <sup>-6</sup>	8343	1:5	-	Positive	Inconclusive	Yes	Experimenter error	XPIVX0122X01 23AGCT082409
CYP	19-May-10	5728	3.26	3.30 x 10 <sup>-6</sup>	6651	1:5	-	Negative	Negative	Yes		2010-05-19XPIVX0116X 0123AgCTT
FAST	16-Nov-09	7966	2.66	2.44 x 10 <sup>-6</sup>	4979	1:5	-	Positive	Inconclusive	No	Failed DMSO control	2009-11-16XPIVAgCTT X0121X0119
FAST	18-Nov-09	3009	3.81	4.26 x 10 <sup>-6</sup>	6145	1:5	-	Negative	Negative	Yes	Experimenter error	2009-11-18 XPIVAgCTX0 120X0121
FAST	07-Dec-09	2286	3.91	2.69 x 10 <sup>-6</sup>	6166	1:5	-	Positive	Inconclusive	Yes		2009-12-07XPIVAgCTX0 107X0121
FEN	12-Nov-09	3889	9.35	4.00 x 10 <sup>-5</sup>	3191	1:2	-	Positive	Positive	Yes	Experimenter error	2009-11-12XPIVAgCTT X0123X0122
FEN	07-Dec-09	2723	4.00	2.00 x 10 <sup>-6</sup>	7034	1:5	-	Positive	Inconclusive	Yes		2009-12-07XPIVAgCTX0 123X0122
FEN	24-Aug-09	2135	3.15	3.26 x 10 <sup>-6</sup>	8343	1:2	-	Positive	Inconclusive	Yes	Experimenter error	XPIVX0122X01 23AGCT082409

Test Substance	Date	DMSO	Induction <sup>1</sup>	E2 Reference Standard EC <sub>50</sub> (µg/mL)	MET	Test Substance Dilution	Test Substance EC <sub>50</sub> (µg/mL)	Positive Using Threshold Method	Positive Using Revised Criteria	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used	Experiment I.D.
FEN	17-May-10	1651	2.04	3.10 x 10 <sup>-6</sup>	6026	1:5	-	Positive	Negative	No	Failed Induction	2010-05-17XPIVAgCTT X0122X0115
FEN	19-May-10	8426	3.28	2.34 x 10 <sup>-6</sup>	7556	1:5	-	Positive	Negative	No	Failed DMSO	2010-05-19XPIVAgCTT X0122X0115
FEN	21-Jun-10	913	3.13	5.19 x 10 <sup>-6</sup>	7505	1:5	1.52 x 10 <sup>0</sup>	Positive	Positive	Yes		2010-06-21X0122X0120 XPIVAgCTRe-run
FLUT	16-Nov-09	7966	2.66	2.44 x 10 <sup>-6</sup>	4979	1:5	-	Positive	Inconclusive	No	Failed DMSO control	2009-11-16XPIVAgCTT X0121X0119
FLUT	18-Nov-09	2453	3.51	3.12 x 10 <sup>-6</sup>	6520	1:5	-	Negative	Negative	Yes	Experimenter error	2009-11-18 XPIVAgCTTX0 119X0111
FLUT	07-Dec-09	2482	4.33	2.80 x 10 <sup>-6</sup>	5868	1:5	-	Positive	Negative	Yes		2009-12-07XPIVAgCTX0 119X0126
FMES	12-Nov-09	4631	5.58	2.52 x 10 <sup>-6</sup>	5148	1:5	-	Positive	Inconclusive	Yes	Experimenter error	2009-11-12XPIVAgCTT X0120X0112
FMES	18-Nov-09	3009	3.81	4.26 x 10 <sup>-6</sup>	6145	1:5	-	Negative	Negative	Yes	Experimenter error	2009-11-18 XPIVAgCTTX0 120X0121
FMES	04-Dec-09	2174	3.72	3.57 x 10 <sup>-6</sup>	4985	1:5	-	Positive	Inconclusive	Yes		2009-12-04XPIVAgCTX0 120X0115
FMES	21-Jun-10	913	3.13	5.19 x 10 <sup>-6</sup>	7505	1:5	7.48 x 10 <sup>0</sup>	Positive	Positive	Yes		2010-06-21X0122X0120 XPIVAgCTRe-run
HPD	16-Nov-09	7302	2.48	2.40 x 10 <sup>-6</sup>	5920	1:5	-	Positive	Negative	No	Failed DMSO and Induction	2009-11-16XPIVAgCTT X0111X0118
HPD	18-Nov-09	2482	3.57	3.96 x 10 <sup>-6</sup>	5932	1:5	-	Positive	Negative	Yes	Experimenter error	2009-11-18 XPIVAgCTTX0 118X0126

ICCVAM BG1Luc ER TA Evaluation Report

Test Substance	Date	DMSO	Induction <sup>1</sup>	E2 Reference Standard EC <sub>50</sub> (µg/mL)	MET	Test Substance Dilution	Test Substance EC <sub>50</sub> (µg/mL)	Positive Using Threshold Method	Positive Using Revised Criteria	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used	Experiment I.D.
HPD	07-Dec-09	2431	4.10	3.24 x 10 <sup>-6</sup>	5577	1:5	-	Positive	Negative	Yes		2009-12-07XPIVAgCTX0109X0118
KCN	28-Dec-09	3675	3.10	3.55 x 10 <sup>-6</sup>	5756	1:5	-	Positive	Negative	Yes		2009-12-28XPIVAgCTT X0117X0127
KCN	13-Nov-09	4753		2.93 x 10 <sup>-6</sup>	2847	1:2	-	Positive	Negative	Yes	Experimenter error	2009-11-13XPIVAgCTT X0125X0117
KCN	07-Dec-09	3541	2.91	1.55 x 10 <sup>-6</sup>	7065	1:5	-	Negative	Negative	No	Failed Induction	2009-12-07XPIVAgCTT X0117X0127
LIN	16-Nov-09	11874	2.61	2.85 x 10 <sup>-6</sup>	4924	1:5	-	Negative	Negative	No	Failed DMSO and Induction	2009-11-16XPIVAgCTT X0107X0116
LIN	13-Dec-09	1810	3.52	2.13 x 10 <sup>-6</sup>	5202	1:5	-	Positive	Inconclusive	Yes		2009-12-13XPIVAgCTT X0160X0116
LIN	17-May-10	1651	2.04	3.10 x 10 <sup>-6</sup>	6026	1:5	-	Negative	Negative	No	Failed Induction	2010-05-17XPIVAgCTT X0122X0115
LIN	19-May-10	8426	3.28	2.34 x 10 <sup>-6</sup>	7556	1:5	-	Negative	Negative	No	Failed DMSO	2010-05-19XPIVAgCTT X0122X0115
LIN	19-May-10	5728	3.26	3.30 x 10 <sup>-6</sup>	6651	1:5	-	Positive	Negative	Yes		2010-05-19XPIVX0116X0123AgCTT
LTX	13-Nov-09	9488	4.32	2.59 x 10 <sup>-6</sup>	-323	1:5	-	Negative	Negative	No	Failed DMSO and MET controls	2009-11-13XPIVAgCTT X0115X0109
LTX	16-Nov-09	5363	3.47	2.10 x 10 <sup>-6</sup>	6300	1:5	-	Positive	Inconclusive	Yes	Experimenter error	2009-11-16XPIVAgCTT X0115X0109
LTX	04-Dec-09	2174	3.72	3.57 x 10 <sup>-6</sup>	4985	1:5	-	Positive	Inconclusive	Yes		2009-12-04XPIVAgCTX0120X0115



Test Substance	Date	DMSO	Induction <sup>1</sup>	E2 Reference Standard EC <sub>50</sub> (µg/mL)	MET	Test Substance Dilution	Test Substance EC <sub>50</sub> (µg/mL)	Positive Using Threshold Method	Positive Using Revised Criteria	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used	Experiment I.D.
LTX	17-Jun-10	1013	3.02	3.50 x 10 <sup>-6</sup>	6621	1:5	-	Positive	Negative	Yes		2010-06-17XPIVX0113X0115AgCTRe-run
MIF	13-Dec-09	1825	3.13	2.74 x 10 <sup>-6</sup>	6247	1:5	-	Positive	Inconclusive	Yes		2009-12-13XPIVAgCTTX0113X0124
MIF	17-Jun-10	1013	3.02	3.50 x 10 <sup>-6</sup>	6621	1:5	-	Positive	Negative	Yes		2010-06-17XPIVX0113X0115AgCTRe-run
MPA	13-Nov-09	7138	4.61	2.45 x 10 <sup>-6</sup>	532	1:2	-	Positive	Negative	No	Failed MET	2009-11-13XPIVAgCTTX0114X0128
MPA	16-Nov-09	8594	3.90	3.08 x 10 <sup>-6</sup>	6466	1:2	-	Negative	Negative	No	Failed DMSO control	2009-11-16XPIVAgCTTX0114X0108
MPA	18-Nov-09	2424	3.92	3.71 x 10 <sup>-6</sup>	5995	1:2	-	Positive	Negative	Yes	Experimenter error	2009-11-18XPIVAgCTTX0114X0108
MPA	13-Dec-09	2100	3.22	2.71 x 10 <sup>-6</sup>	5875	1:2	-	Positive	Inconclusive	Yes		2009-12-13XPIVAgCTTX0108X0114
MPA	21-Jun-10	865	3.40	3.50 x 10 <sup>-6</sup>	6262	1:2	-	Positive	Negative	Yes		2010-06-21X0114XPIVAgCTRe-run
NIL	12-Nov-09	4631	5.58	2.52 x 10 <sup>-6</sup>	5148	1:5	-	Positive	Inconclusive	Yes	Experimenter error	2009-11-12XPIVAgCTTX0120X0112
NIL	04-Dec-09	2187	3.93	3.87 x 10 <sup>-6</sup>	3761	1:5	-	Positive	Positive	Yes		2009-12-04XPIVAgCTTX0112X0125
NORT	16-Nov-09	5277	3.43	1.24 x 10 <sup>-6</sup>	6378	1:5	4.93 x 10 <sup>-1</sup>	Positive	Positive	Yes		2009-11-16XPIVAgCTTX0127X0130
OHAN	13-Dec-09	1382	4.42	1.82 x 10 <sup>-6</sup>	4973	1:2	1.62 x 10 <sup>+1</sup>	Positive	Positive	Yes		2009-12-13XPIVAgCTTX0129X0110

ICCVAM BG1Luc ER TA Evaluation Report

Test Substance	Date	DMSO	Induction <sup>1</sup>	E2 Reference Standard EC <sub>50</sub> (µg/mL)	MET	Test Substance Dilution	Test Substance EC <sub>50</sub> (µg/mL)	Positive Using Threshold Method	Positive Using Revised Criteria	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used	Experiment I.D.
OHAN	21-Jun-10	915	3.43	1.59 x 10 <sup>-5</sup>	5108	1:2	7.46 x 10 <sup>0</sup>	Positive	Positive	Yes		2010-06-21X0129X0160 XPIVAgCTRe-Run
OX	16-Nov-09	7302	2.48	2.40 x 10 <sup>-6</sup>	5920	1:2	-	Positive	Negative	No	Failed DMSO and Induction	2009-11-16XPIVAgCTT X0111X0118
OX	18-Nov-09	2453	3.51	3.12 x 10 <sup>-6</sup>	6520	1:2	-	Positive	Negative	Yes	Experimenter error	2009-11-18 XPIVAgCTTX0 119X0111
OX	07-Dec-09	2386	3.92	2.68 x 10 <sup>-6</sup>	5849	1:2	-	Positive	Negative	Yes		2009-12-07XPIVAgCTX0 111X0128
PCY	13-Nov-09	9488	4.32	2.59 x 10 <sup>-6</sup>	-323	1:5	-	Negative	Negative	No	Failed DMSO and MET controls	2009-11-13XPIVAgCTT X0115X0109
PCY	16-Nov-09	5363	3.47	2.10 x 10 <sup>-6</sup>	6300	1:5	-	Negative	Negative	Yes	Experimenter error	2009-11-16XPIVAgCTT X0115X0109
PCY	07-Dec-09	2431	4.10	3.24 x 10 <sup>-6</sup>	5577	1:5	-	Positive	Negative	Yes		2009-12-07XPIVAgCTX0 109X0118
PZE	13-Dec-09	1382	4.42	1.82 x 10 <sup>-6</sup>	4973	1:2	-	Positive	Inconclusive	Yes		2009-12-13XPIVAgCTT X0129X0110
RSP	12-Nov-09	7749	5.46	2.50 x 10 <sup>-6</sup>	4458	1:2	-	Positive	Inconclusive	No	Failed DMSO control	2009-11-12XPIII-IVAgCTTX0065 X0108
RSP	16-Nov-09	8594	3.90	3.08 x 10 <sup>-6</sup>	6466	1:2	-	Negative	Negative	No	Failed DMSO control	2009-11-16XPIVAgCTT X0114X0108
RSP	18-Nov-09	2424	3.92	3.71 x 10 <sup>-6</sup>	5995	1:2	-	Positive	Negative	Yes	Experimenter error	2009-11-18 XPIVAgCTTX0 114X0108
RSP	13-Dec-09	2100	3.22	2.71 x 10 <sup>-6</sup>	5875	1:2	-	Positive	Inconclusive	Yes		2009-12-13XPIVAgCTT X0108X0114

Test Substance	Date	DMSO	Induction <sup>1</sup>	E2 Reference Standard EC <sub>50</sub> (µg/mL)	MET	Test Substance Dilution	Test Substance EC <sub>50</sub> (µg/mL)	Positive Using Threshold Method	Positive Using Revised Criteria	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used	Experiment I.D.
RSP	19-May-10	5987	3.00	2.80 x 10 <sup>-6</sup>	8843	1:5	-	Positive	Negative	Yes		2010-05-19XPIVAgCTT X0108X0131
SPIR	16-Nov-09	11874	2.61	2.85 x 10 <sup>-6</sup>	4924	1:5	-	Negative	Negative	No	Failed DMSO and Induction	2009-11-06XPIVAgCTT X0107X0116
SPIR	07-Dec-09	2286	3.91	2.69 x 10 <sup>-6</sup>	6166	1:5	-	Positive	Negative	Yes		2009-12-07XPIVAgCTX0107X0121
TREN	13-Dec-09	1488	4.02	1.95 x 10 <sup>-6</sup>	6076	1:2	-	Positive	Inconclusive	Yes		2009-12-13XPIVAgCTT X0131X0150
TREN	19-May-10	5987	3.00	2.80 x 10 <sup>-6</sup>	8843	1:5	2.59 x 10 <sup>-2</sup>	Positive	Positive	Yes		2010-05-19XPIVAgCTT X0108X0131

Abbreviations: AMP = Ammonium perchlorate; APO = Apomorphine; BENZ = 2,2',4,4'-Tetrahydroxybenzophenone; BICAL = Bicalutamide; CHX = Cycloheximide; CHY = Chrysin; CYP = Cyproterone acetate; DMSO = Dimethyl sulfoxide; EC<sub>50</sub> = half maximal effective concentration; E2 = 17β-estradiol; FAST = Finasteride; FEN = Fenarimol; FLUT = Flutamide; FMES = Fluoxymestron; HPD = Haloperidol; KCN = Ketoconazole; LIN = Linuron; LTX = L-thyroxine; MET = *p,p'*-methoxychlor; MIF = Mifepristone; MPA = Medroxyprogesterone acetate; NIL = Nilutamide; NORT = 19-nortestosterone; OHAN = 4-hydroxyandrostenedione; OX = Oxazepam; PCY = Procymidone; PZE = Pimozide; RSP = Reserpine; SPIR = Spironolactone; TREN = 17β-trenbolone

\* Benz (2,2',4,4'-Tetrahydroxybenzophenone, CASRN 131-55-5) was tested once at XDS as part of a cooperative effort with the OECD performance-based test guidelines working group.

<sup>1</sup> Induction for comprehensive test plates is measured by dividing the averaged highest E2 reference standard RLU value by the averaged DMSO control RLU value.

**Table 4 Phase IV Antagonist Plates Tested at XDS**

Test Substance	Date	DMSO	Reduction <sup>1</sup>	Ral/E2 Reference Standard IC <sub>50</sub> (µg/mL)	E2 Cont	Fla\E2 Cont	Test Substance Dilution	Test Substance IC <sub>50</sub> (µg/mL)	Positive Using Threshold Method	Positive Using Revised Criteria	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used	Experiment I.D.
AMP	10-Dec-09	3153	9.07	3.13 x 10 <sup>-4</sup>	7724	869	1:2	-	Positive	Negative	Yes		2009-12-10XPIVANTCTX0147X0153
APO	30-Nov-09	3162	4.74	2.61 x 10 <sup>-4</sup>	8232	3540	1:2	-	Positive	Positive	Yes		2009-11-30XPIIIAntCTX0152X0105
BICAL	18-Dec-09	4300	9.22	2.98 x 10 <sup>-4</sup>	7138	965	1:2	1.99 x 10 <sup>+1</sup>	Positive	Negative	Yes		2009-12-10XPIVANTCTX0151X0144
CHX	21-Dec-09	3755	8.69	2.95 x 10 <sup>-4</sup>	6748	1457	1:5	2.72 x 10 <sup>-1</sup>	Positive	Positive	Yes		2009-12-10XPIVAntCTX0139X0149
CHY	15-Dec-09	4615	5.74	3.37 x 10 <sup>-4</sup>	8279	1209	1:2	-	Negative	Negative	Yes		2009-12-15XPIVANTCTX0135X0150
CYP	18-Dec-09	3612	10.29	3.52 x 10 <sup>-4</sup>	7097	871	1:2	1.98 x 10 <sup>+1</sup>	Positive	Negative	Yes		2009-12-10XPIVANTCTX0141X0148
FAST	15-Dec-09	4737	6.20	4.29 x 10 <sup>-4</sup>	5662	430	1:2	-	Positive	Negative	No	Failed E2 Control	2009-12-15XPIVANTCTX0146X0142
FAST	15-Dec-09	2631	6.10	2.35 x 10 <sup>-4</sup>	8658	3071	1:2	4.72 x 10 <sup>+1</sup>	Positive	Negative	Yes		2009-12-15XPIVANTCTX0146
FEN	19-Aug-09	3691	5.22	8.13 x 10 <sup>-4</sup>	4406	1344	1:2	-	Positive	Negative	No	Failed E2 Control	XPIII0067X0147ANTCT081909
FEN	04-Nov-09	10979	5.48	3.20 x 10 <sup>-4</sup>	7854	2330	1:2	-	Negative	Negative	No	Failed DMSO Control	2009-11-04AntCTX0082X0147
FEN	5-Nov-09	8712	5.93	2.97 x 10 <sup>-4</sup>	7691	2494	1:2	-	Negative	Negative	No	Failed DMSO control	2009-11-05X0082X0147Ant
FEN	10-Dec-09	3153	9.07	3.13 x 10 <sup>-4</sup>	7724	869	1:2	4.40 x 10 <sup>+2</sup>	Positive	Negative	Yes		2009-12-10XPIVANTCTX0147X0153

Test Substance	Date	DMSO	Reduction <sup>1</sup>	Ral/E2 Reference Standard IC <sub>50</sub> (µg/mL)	E2 Cont	Fla/E2 Cont	Test Substance Dilution	Test Substance IC <sub>50</sub> (µg/mL)	Positive Using Threshold Method	Positive Using Revised Criteria	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used	Experiment I.D.
FLUT	18-Dec-09	4300	9.22	2.98 x 10 <sup>-4</sup>	7138	965	1:5	4.33 x 10 <sup>0</sup>	Positive	Negative	Yes		2009-12-10XPIVANTCTX0151X0144
FMES	10-Dec-09	3299	11.20	3.01 x 10 <sup>-4</sup>	6519	755	1:2	-	Negative	Negative	Yes		2009-12-10XPIVANTCTX0138X0145
HPD	15-Dec-09	5539	5.37	4.35 x 10 <sup>-4</sup>	8064	1223	1:2	3.89 x 10 <sup>0</sup>	Positive	Negative	Yes		2009-12-15XPIVANTCTX0154X0143
KCN	21-Dec-09	3194	4.81	1.56 x 10 <sup>-4</sup>	6097	2024	1:2	-	Negative	Negative	No	Failed E2, Fla/E2 Control	2009-12-15XPIVANTCTX0134X0142
KCN	15-Dec-09	4737	6.20	4.29 x 10 <sup>-4</sup>	5662	430	1:2	-	Positive	Negative	No	Failed E2 Control	2009-12-15XPIVANTCTX0146X0142
KCN	22-Dec-09	3111	4.23	2.66 x 10 <sup>-4</sup>	7091	1892	1:2	6.55 x 10 <sup>-1</sup>	Positive	Positive	Yes		2009-12-22XPIVANTCTX0134X0142
LIN	18-Dec-09	3612	10.29	3.52 x 10 <sup>-4</sup>	7097	871	1:2	3.34 x 10 <sup>-1</sup>	Positive	Negative	Yes		2009-12-10XPIVANTCTX0141X0148
LTX	21-Dec-09	2591	5.38	2.41 x 10 <sup>-4</sup>	7889	1254	1:2	-	Positive	Negative	Yes		2009-12-15XPIVAntCTX0132X0140
MIF	10-Dec-09	3299	11.20	3.01 x 10 <sup>-4</sup>	6519	755	1:2	8.21 x 10 <sup>-1</sup>	Positive	Negative	Yes		2009-12-10XPIVANTCTX0138X0145
MPA	21-Dec-09	3755	8.69	2.95 x 10 <sup>-4</sup>	6748	1457	1:2	-	Positive	Positive	Yes		2009-12-10XPIVAntCTX0139X0149
NIL	21-Dec-09	2956	5.46	2.73 x 10 <sup>-4</sup>	6166	1081	1:5	-	Positive	Negative	No	Failed E2 Control	2009-12-15XPIVANTCTX0134X0137
NIL	15-Dec-09	3130	4.66	2.19 x 10 <sup>-4</sup>	7180	1818	1:5	9.28 x 10 <sup>0</sup>	Positive	Negative	Yes		2009-12-15XPIVANTCTX0156X0137

ICCVAM BG1Luc ER TA Evaluation Report

Test Substance	Date	DMSO	Reduction <sup>1</sup>	Ral/E2 Reference Standard IC <sub>50</sub> (µg/mL)	E2 Cont	Fla/E2 Cont	Test Substance Dilution	Test Substance IC <sub>50</sub> (µg/mL)	Positive Using Threshold Method	Positive Using Revised Criteria	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used	Experiment I.D.
NIL	28-Dec-09	3500	6.07	2.15 x 10 <sup>-4</sup>	6207	1456	1:5	2.38x 10 <sup>+1</sup>	Positive	Negative	No	Failed E2 Control	2009-12-28-XPIVANTCTX0156X0137
NIL	29-Dec-09	6482	6.22	3.21 x 10 <sup>-4</sup>	6598	166	1:5	1.23 x 10 <sup>+1</sup>	Positive	Negative	Yes		2009-12-29-XPIVANTCTX0156X0137
NORT	10-Dec-09	4250	7.97	3.37 x 10 <sup>-4</sup>	7106	1393	1:2	5.02 x 10 <sup>+1</sup>	Positive	Negative	Yes		2009-12-10XPIIVANTCT X0155X0019
OHAN	15-Dec-09	5539	5.37	4.35 x 10 <sup>-4</sup>	8064	1223	1:2	-	Negative	Negative	Yes		2009-12-15XPIVANTCTX 0154X0143
OX	21-Dec-09	2498	5.39	1.97 x 10 <sup>-4</sup>	7102	1173	1:2	3.30 x 10 <sup>+1</sup>	Positive	Negative	Yes		2009-12-15XPIVANTCTX 0133X0136
PCY	21-Dec-09	2956	5.46	2.73 x 10 <sup>-4</sup>	6166	1081	1:2	-	Negative	Negative	No	Failed E2 Control	2009-12-15XPIVANTCTX 0134X0137
PCY	21-Dec-09	3194	4.81	1.56 x 10 <sup>-4</sup>	6097	2024	1:2	-	Negative	Negative	No	Failed E2, Fla/E2 Control	2009-12-15XPIVANTCTX 0134X0142
PCY	22-Dec-09	3111	4.23	2.66 x 10 <sup>-4</sup>	7091	1892	1:2	-	Negative	Negative	Yes		2009-12-22XPIVANTCTX 0134X0142
PMZ	15-Dec-09	4615	5.74	3.37 x 10 <sup>-4</sup>	8279	1209	1:2	-	Positive	Negative	Yes		2009-12-15XPIVANTCTX 0135X0150
RSP	21-Dec-09	2498	5.39	1.97 x 10 <sup>-4</sup>	7102	1173	1:2	-	Positive	Negative	Yes		2009-12-15XPIVANTCTX 0133X0136
SPIR	21-Dec-09	2591	5.38	2.41 x 10 <sup>-4</sup>	7889	1254	1:2	-	Positive	Negative	Yes		2009-12-15XPIVAntCTX0 132X0140
TREN	15-Dec-09	744	2.77	1.04 x 10 <sup>-3</sup>	4172	222	1:2	-	Positive	Negative	No	Failed E2 Control, Reduction	2009-12-15XPIVANTCTX 0156

Test Substance	Date	DMSO	Reduction <sup>1</sup>	Ral/E2 Reference Standard IC <sub>50</sub> (µg/mL)	E2 Cont	Fla/E2 Cont	Test Substance Dilution	Test Substance IC <sub>50</sub> (µg/mL)	Positive Using Threshold Method	Positive Using Revised Criteria	Was Plate Used for Data Analysis?	Reason Why Plate Was Not Used	Experiment I.D.
TREN	15-Dec-09	3130	4.66	2.19 x 10 <sup>-4</sup>	7180	1818	1:2	-	Positive	Negative	Yes		2009-12-15XPIVANTCTX0156X0137
TREN	28-Dec-09	3500	6.07	2.15 x 10 <sup>-4</sup>	6207	1456	1:2	-	Positive	Negative	No	Failed E2 Control	2009-12-28-XPIVANTCTX0156X0137
TREN	29-Dec-09	6482	6.22	3.21 x 10 <sup>-4</sup>	6598	166	1:2	-	Positive	Negative	Yes		2009-12-29-XPIVANTCTX0156X0137

Abbreviations: AMP = Ammonium perchlorate; APO = Apomorphine; BICAL = Bicalutamide; CHX = Cycloheximide; CHY = Chrysin; Cont = Control; CYP = Cyproterone acetate; DMSO = Dimethyl sulfoxide; E2 = 17β-estradiol; FAST = Finasteride; FEN = Fenarimol; Fla = Flavone; FLUT = Flutamide; FMES = Fluoxymestron; HPD = Haloperidol; IC<sub>50</sub> = half-maximal inhibitory concentration; I.D. = Identification; KCN = Ketoconazole; LIN = Linuron; LTX = L-thyroxine; MET = *p,p'*-methoxychlor; MIF = Mifepristone; MPA = Medroxyprogesterone acetate; NIL = Nilutamide; NORT = 19-nortestosterone; OHAN = 4-hydroxyandrostenedione; OX = Oxazepam; PCY = Procymidone; PZE = Pimozide; Ral = Raloxifene HCl; RSP = Reserpine; SPIR = Spironolactone; TREN = 17β-trenbolone

<sup>1</sup> Reduction for comprehensive test plates is measured by dividing the averaged highest Ral/E2 reference standard RLU value by the lowest averaged Ral/E2 reference standard RLU value.